Q Logic 2 Manual Updating Modules









Replacing AM2 with AAM

The AAM is installed when 3 or more power function are on unit (lift, tilt & recline).








Identify the Symptom

Error #251 A device with 'Incompatible Firmware Version' is connected to the system





What do the symptoms indicate?

A #251 error indicates a module connected in the system has firmware installed which is incompatible with the other modules.

What Parts could cause symptom?

Any module connected in the Bus System.



Incompatible Firmware Version

✓ Using the Hand Held or PC programmers to check the software versions of each component.



Update the firmware on the chair so all components match

- What do we update?
 - ✓ If the firmware for the majority of components are Pre 2012 firmware and the mismatched component is 2012 Firmware


 - ✓ Or update pre 2012 components to 2012 firmware
 - Making sure the patient is ok with any changes to the system
 that may come with new firmware (i.e. Sip N Puff gauge on screen)
 - ✓ If the firmware for majority of components is 2012 Firmware and the mismatched component is pre 2012 Firmware.
 - ✓ Update the pre 2012 component to the 2012 Firmware



Check firmware levels with HHP.

Verify needed firmware versions are installed.














First Verify Auto Shutoff has been turned off!



































































Manually Updating Modules iLevel Program(CPF) file must be installed











































Thank you for your time

Quantum Rehab Clinical Education Department

education@pridemobility.com

